

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produ	ct: AM304F			
		DE STANDARD 50 PPM IN TO	OLUENE	
Test		Target/UOM	Range	Result
PREPARED TO FOR	MULATION ON	YES	G	YES
FILE		YES		
Sample	ID			
Tested	10/25/24			
Tester	JTRA			
Recorde	d 10/25/24			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample	ID	-		
Tested	10/25/24			
Tester	JTRA			
Recorde				
INSTRUMENTS USEI		INSTRUMENT		P400,P397,VF200-T18
PREPARATION	Donard			
Sample	חו			
Tested	10/25/24			
Tester	JTRA			
Recorde				
PPM CHLORIDE	u 10/25/24	50.00	49.50 - 50.50	49.99
		PPM	49.30 - 30.30	49.99
		PPM		
Sample				
Tested	10/25/24			
Tester	JTRA			
Recorde	d 10/25/24			10/00/00
EXPIRATION DATE		MM/DD/YY		10/30/26
		MM/DD/YYYY		
Sample				
Tested	10/25/24			
Tester	JTRA			
Recorde	d 10/25/24			
Lot # 410	2575			
Made 10/	25/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1