



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SPE509F			
TOTAL SUSPENDED SOLIDS STANDARD 1000 PPM			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 10/28/24			
Tester AROS			
Recorded 10/28/24			
APPEARANCE AND COLOR	TURBID LIQUID APPEARANCE		PASS
Sample ID			
Tested 10/28/24			
Tester AROS			
Recorded 10/28/24			
CALCULATED MG/L (PPM) TSS	1000.0 MG/L	995.0 - 1005.0	1000.3
Sample ID			
Tested 10/28/24			
Tester AROS			
Recorded 10/28/24			
ACT GMS CELITE	1.0000 GM	0.9990 - 1.0010	1.0003
Sample ID			
Tested 10/28/24			
Tester AROS			
Recorded 10/28/24			
QCSP11:STANDARDIZATION OF BROMIDE-BROMAT	YES QCSP11		YES
Sample ID			
Tested 10/28/24			
Tester AROS			
Recorded 10/28/24			
MLS OF WATER	1000 ML	995 - 1005	1000
Sample ID			
Tested 10/28/24			
Tester AROS			
Recorded 10/28/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M108,VF1-Z39
Sample ID			
Tested 10/28/24			
Tester AROS			
Recorded 10/28/24			
WEIGHT NIST TRACEABLE	PASS YES		PASS
Sample ID			
Tested 10/28/24			
Tester AROS			
Recorded 10/28/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

SHELF LIFE OF PRODUCT	SHELF LIFE TIME	1 YEAR
Sample ID		
Tested 10/28/24		
Tester AROS		
Recorded 10/28/24		
EXPIRATION DATE	MM/DD/YYYY	10/30/2025
	MM/DD/YYYY	
Sample ID		
Tested 10/28/24		
Tester AROS		
Recorded 10/28/24		
Lot # 4102819		
Made 10/28/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager