

Page: 1

Date: 10/21/24 at 2:11 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc				
-	FLUORIDE	E STD 1,000 PPM W/V CE		B
Test		Target/UOM	Range	Result
PREPARED TO FORM	ULATION ON	YES		YES
FILE		YES		
Sample II				
Tested	10/21/24			
Tester	RIRO			
Recorded				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample II)			
Tested	10/21/24			
Tester	RIRO			
Recorded	10/21/24			
INSTRUMENTS USED DURING		INSTRUMENT		M118,VF3-J18,BOTTLE
PREPARATION				
Sample II)			
Tested	10/21/24			
Tester	RIRO			
Recorded				
PPM FLUORIDE		1000	980 - 1020	999
		PPM		
Sample II)			
Tested	10/21/24			
Tester	RIRO			
Recorded				
QCSP45:MEASURING		YES		YES
WATER		QCSP45		120
Sample II	`	Q001 40		
Tested				
	10/21/24			
Tester	RIRO			
Recorded	10/21/24	MM/DD/WV		04/20/25
EXPIRATION DATE		MM/DD/YY		04/30/25
=		MM/DD/YYYY		
Sample II				
Tested	10/21/24			
Tester	RIRO			
Recorded				
SHELF LIFE OF PROD	OUCT	6 MONTHS		6 MONTHS
0 : :-		TIME		
Sample II				
Tested	10/21/24			
Tester	RIRO			
Recorded				
Lot # 4102				
Made 10/2	1/24			



Page: 2

Date: 10/21/24 at 2:11 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager