

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Packaged	Product:	6919			
ruonugeu			IVITY STANDARD 100 uMHOS		
Test			Target/UOM	Range	Result
	TO FORMULA		YES	nange	YES
FILE			YES		-
	Sample ID				
		0/24/24			
		PHI			
		0/24/24			
CONDUCTIV			100	99 - 101	100
	-		US/CM		
	Sample ID				
	-	0/24/24			
		PHI			
		0/24/24			
QCSP30:CO	NDUCTIVITY		YES		YES
STD&MEAS	UREMENT		QCSP30		
	Sample ID				
	•	0/24/24			
		PHI			
		0/24/24			
TRACEABLE	E TO N.I.S.T. SI	RM 3191	YES		YES
			YES		
	Sample ID				
	-	0/24/24			
		PHI			
	Recorded 10	0/24/24			
APPEARAN	CE AND COLO	R	CLEAR WATER WHITE		PASS
			LIQUID		
			APPEARANCE		
	Sample ID				
	-	0/24/24			
		PHI			
	Recorded 10	0/24/24			
INSTRUMEN	ITS USED DUR	RING	INSTRUMENT		M92,T-229,BOTTLE
PREPARATI	ON				
	Sample ID				
		0/24/24			
	Tester B	PHI			
	Recorded 10	0/24/24			
SHELF LIFE	OF PRODUCT	ſ	1 YEAR		1 YEAR
			TIME		
	Sample ID				
	Tested 10	0/24/24			
	Tester B	PHI			
		0/24/24			
EXPIRATION	I DATE		MM/DD/YY		10/30/25
			MM/DD/YYYY		
	Sample ID				
		0/24/24			
		PHI			
	Recorded 10	0/24/24			



## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4102449
Made	10/24/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

thand St

**Quality Manager**