10/25/24

Made

Page: 1

Date: 10/25/24 at 1:59 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc		E STANDARD 100 PPM AS	CL2 CERTIFIED SOLUTION	
Test	CILIDICIN	Target/UOM	Range	Result
Actual Density		0.99951	0.99893 - 0.99968	0.99946
Actual Delisity		G/ML	0.99090 - 0.99900	0.225-0
Sample I)	<u> </u>		
Tested	10/25/24			
Tester	JENC			
Recorded				
INSTRUMENTS USED		INSTRUMENT		M112,VF12-5,M55
PREPARATION				, ,
Sample I)			
Tested	10/25/24			
Tester	JENC			
Recorded				
EXPIRATION DATE		MM/DD/YY		12/30/24
		MM/DD/YYYY		
Sample II)			
Tested	10/25/24			
Tester	JENC			
Recorded	10/25/24			
SHELF LIFE OF PRODUCT		2 MONTHS		2 MONTHS
		TIME		
Sample II)			
Tested	10/25/24			
Tester	JENC			
Recorded	10/25/24			
PPM CHLORIDE		100.00	98.00 - 102.00	100.80
		PPM		
Sample II)			
Tested	10/25/24			
Tester	JENC			
Recorded	10/25/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Proposition of the