



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: I7604-4L ISOPROPYL ALCOHOL HPLC GRADE (PX1834-1)			
Test	Target/UOM	Range	Result
ASSAY OF MATERIAL	MIN 99.5% %		>99.99%
Sample ID			
Tested	10/22/24		
Tester	TBUR		
Recorded	10/22/24		
WATER CONTENT	MAX 0.01% %		0.0012%
Sample ID			
Tested	10/22/24		
Tester	TBUR		
Recorded	10/22/24		
UV CUTOFF	204 NM NM		202.3 NM
Sample ID			
Tested	10/22/24		
Tester	TBUR		
Recorded	10/22/24		
REFRACTIVE INDEX	1.3700 REFRACTIVE INDE	0.0000 - 1.3800	1.3771
Sample ID			
Tested	10/22/24		
Tester	TBUR		
Recorded	10/22/24		
ELECTRON CAPTURE	PASSES TEST ELECTRON C		PASSES TEST
Sample ID			
Tested	10/22/24		
Tester	TBUR		
Recorded	10/22/24		
RESIDUE AFTER EVAPORATION	MAX 1 PPM		<0.5 PPM
Sample ID			
Tested	10/22/24		
Tester	TBUR		
Recorded	10/22/24		
MANUFACTURED DATE OF MATERIAL	MM/DD/YYYY MM/DD/YYYY		10/07/2024
Sample ID			
Tested	10/22/24		
Tester	TBUR		
Recorded	10/22/24		
Lot #	429507		
Made	10/21/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 10/22/24 at 12:04 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager