



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SYN034			
AQUALINE COMPLETE 2(1 LITER) *DEA Regulated			
Test	Target/UOM	Range	Result
APPEARANCE	REPORT		BROWN MOBILE LIQUID
Sample ID			
Tested	10/29/24		
Tester	TBUR		
Recorded	10/29/24		
EQUIVALENCE IN MG/ML	>= 2.2 MG/ML		2.43
Sample ID			
Tested	10/29/24		
Tester	TBUR		
Recorded	10/29/24		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		08/30/2025
Sample ID			
Tested	10/29/24		
Tester	TBUR		
Recorded	10/29/24		
Lot #	430205		
Made	10/28/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager