

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Packaged F	Product:	UCT126F			
		CELITE	STD FOR TSS 40 MG/L SOLU	JTION	
Test			Target/UOM	Range	Result
PREPARED T	O FORMUL	ATION ON	YES		YES
FILE			YES		
S	ample ID				
Т	ested	11/06/24			
Т	ester	AROS			
F	Recorded	11/06/24			
ACTUAL PPM			40.00	39.50 - 40.50	40.08
			PPM		
s	ample ID				
Т	ested	11/06/24			
Т	ester	AROS			
F	Recorded	11/06/24			
APPEARANCE AND COLOR		.OR	SLIGHTLY TURBID LIQUID		PASS
			APPEARANCE		
S	ample ID				
Т	ested	11/06/24			
Т	ester	AROS			
F	Recorded	11/06/24			
INSTRUMENTS USED DURING		URING	INSTRUMENT		M108,VF4-N32
PREPARATIO	N				
s	ample ID				
		11/06/24			
Т	ester	AROS			
F	Recorded	11/06/24			
EXPIRATION DATE			MM/DD/YY		05/30/25
			MM/DD/YYYY		
s	ample ID				
		11/06/24			
Т	ester	AROS			
F	Recorded	11/06/24			
Lot #	411061	12			
Made	11/06,				
waue	11,007	, _ 1			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA