

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	CG047F			
Ŭ		itrogen 1000 mg/L Labo	ratory Control Sample	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES	<u>v</u>	YES
FILE		YES		
Sample ID				
Tested	11/07/24			
Tester	AROS			
Recorded	11/07/24			
Calculated ppm Nitrog	en	1000.0	995.0 - 1005.0	999.9
		PPM		
Sample ID				
Tested	11/07/24			
Tester	AROS			
Recorded	11/07/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID		-		
Tested	11/07/24			
Tester	AROS			
Recorded	11/07/24			
INSTRUMENTS USED		INSTRUMENT		M108,GC,VF1-Z16
PREPARATION				
Sample ID				
Tested	11/07/24			
Tester	AROS			
Recorded	11/07/24			
EXPIRATION DATE	11/07/24	MM/DD/YYYY		02/07/2025
EXPIRATION DATE		MM/DD/YYYY		02/01/2023
Sample ID				
Tested	11/07/24			
Tester	AROS			
Recorded	11/07/24			
Lot # 4110	770			
Made 11/0	7/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA