



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: ND497F			
Density Standard Ultra Pure Water			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 11/08/24			
Tester JTRA			
Recorded 11/08/24			
DENSITY @ 10°C	0.99970 GM/CM3	0.99968 - 0.99972	0.99970
Sample ID			
Tested 11/08/24			
Tester JTRA			
Recorded 11/08/24			
DENSITY @ 15°C	0.99910 GM/CM3	0.99908 - 0.99912	0.99910
Sample ID			
Tested 11/08/24			
Tester JTRA			
Recorded 11/08/24			
DENSITY @ 15.56°C (60°F)	0.99901 GM/ML	0.99899 - 0.99903	0.99901
Sample ID			
Tested 11/08/24			
Tester JTRA			
Recorded 11/08/24			
DENSITY @ 25°C	0.99704 GM/CM3	0.99702 - 0.99706	0.99704
Sample ID			
Tested 11/08/24			
Tester JTRA			
Recorded 11/08/24			
DENSITY @ 30°C	0.99565 GM/CM3	0.99563 - 0.99567	0.99565
Sample ID			
Tested 11/08/24			
Tester JTRA			
Recorded 11/08/24			
DENSITY @ 35°C	0.99403 GM/CM3	0.99401 - 0.99405	0.99403
Sample ID			
Tested 11/08/24			
Tester MMAT			
Recorded 11/08/24			
DENSITY @ 40°C	0.99221 GM/CM3	0.99219 - 0.99223	0.99221
Sample ID			
Tested 11/08/24			
Tester JTRA			
Recorded 11/08/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

DENSITY @ 45°C	0.99021 GM/CM3	0.99019 - 0.99023	0.99021
Sample ID			
Tested	11/08/24		
Tester	JTRA		
Recorded	11/08/24		
DENSITY @ 50°C	0.98803 GM/CM3	0.98801 - 0.99805	0.98803
Sample ID			
Tested	11/08/24		
Tester	JTRA		
Recorded	11/08/24		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	11/08/24		
Tester	JTRA		
Recorded	11/08/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		05/30/26
Sample ID			
Tested	11/08/24		
Tester	JTRA		
Recorded	11/08/24		
DENSITY @20°C	0.99820 GM/CM3	0.99818 - 0.99822	0.99820
Sample ID			
Tested	11/08/24		
Tester	JTRA		
Recorded	11/08/24		
Lot #	4110703		
Made	11/08/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager