



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 3650			
EDTA TITRANT 0.0575 MOLAR (0.115N)			
Test	Target/UOM	Range	Result
SHELF LIFE OF PRODUCT	2 YEARS		2 YEARS
	TIME		
Sample ID			
Tested	11/15/24		
Tester	ENOR		
Recorded	11/15/24		
PREPARED TO FORMULATION ON FILE	YES		YES
	YES		
Sample ID			
Tested	11/15/24		
Tester	ENOR		
Recorded	11/15/24		
Actual Molarity	0.0575	0.0570 - 0.0580	0.0578
	MOL/L		
Sample ID			
Tested	11/15/24		
Tester	ENOR		
Recorded	11/15/24		
QCSP4:STANDARDIZATION EDTA SOLUTIONS	YES		YES
	QCSP4		
Sample ID			
Tested	11/15/24		
Tester	ENOR		
Recorded	11/15/24		
N.I.S.T. Traceable to SRM 915	YES		YES
	YES		
Sample ID			
Tested	11/15/24		
Tester	ENOR		
Recorded	11/15/24		
Actual Density	1.00888	1.00837 - 1.00939	1.00903
	G/ML		
Sample ID			
Tested	11/15/24		
Tester	ENOR		
Recorded	11/15/24		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	11/15/24		
Tester	ENOR		
Recorded	11/15/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M112,T220,M55
Sample ID			
Tested	11/15/24		
Tester	ENOR		
Recorded	11/15/24		



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

EXPIRATION DATE	MM/DD/YY	11/30/26
	MM/DD/YYYY	
Sample ID		
Tested	11/15/24	
Tester	ENOR	
Recorded	11/15/24	
Lot #	4110437	
Made	11/15/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager