

Recorded 11/13/24

Page: 1

Date: 11/14/24 at 7:02 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # ·

Customer PO # :						
Packaged Product:	8430					
	SILVER N	IITRATE 0.1 NORMAL NIS	ST TRACEABLE SOLUTION			
Test		Target/UOM	Range	Result		
PREPARED TO FORMUL	ATION ON	YES		YES		
FILE		YES				
Sample ID						
	11/13/24					
Tester	RIRO					
	11/13/24					
SAMPLE COLLECTION & HANDLING		YES		YES		
PER LAB040		YES				
Sample ID						
	11/13/24					
	RIRO					
	11/13/24					
DATE OF SAMPLING FOR		MM/DD/YYYY		11/13/2024		
PURPOSES		MM/DD/YYYY				
Sample ID		, 22,				
· ·	11/13/24					
	RIRO					
	11/13/24					
AMBIENT TEMPERATUR		25.0	10.0 - 35.0	25.0		
SAMPLING DEGC	L DOMINO	DEG C	10.0 00.0	20.0		
Sample ID		<i>DEG</i>				
· ·	11/12/24					
	11/13/24 RIRO					
	11/13/24					
DATE OF TESTING OF SA		MM/DD/YYYY		11/13/2024		
	AWIPLE	MM/DD/YYYY		11/13/2024		
Sample ID						
Tested	11/13/24					
	RIRO					
Recorded	11/13/24					
AMBIENT TEST TEMPER	ATURE	20.0	15.0 - 30.0	19.0		
DEG C		DEG C				
Sample ID						
	11/13/24					
	RIRO					
	11/13/24					
Actual Normality		0.1000	0.0995 - 0.1005	0.1004		
		EQ/L				
Sample ID						
Tested	11/13/24					
Tester	RIRO					
Recorded	11/13/24					
QCSP53: SILVER NITRATE STD ON		YES		YES		
AUTO-TITRA		METHOD				
Sample ID						
•	11/13/24					
Tester	RIRO					



Made 11/13/24

Page: 2

Date: 11/14/24 at 7:02 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

		Cusio	mer PO#.	
Actual Density		1.01238	1.01192 - 1.01284	1.01245
		G/ML		
Sample ID		5,		
	11/13/24			
	RIRO			
	11/13/24			
VERIFICATION BY QCSP		PASS		PASS
VERIFICATION DT QCSP	70	PASS		FA55
0 1 10		PASS		
Sample ID				
	11/13/24			
	RIRO			
	11/13/24			
N.I.S.T. Traceable to SRM 919 NaCl		YES		YES
		YES		
Sample ID				
Tested	11/13/24			
Tester	RIRO			
Recorded	11/13/24			
APPEARANCE AND COLOR		CLEAR WATER WHITE		PASS
		LIQUID		
		APPEARANCE		
Sample ID				
	11/13/24			
	RIRO			
	11/13/24			
INSTRUMENTS USED DU		INSTRUMENT		M113,T261,M55
PREPARATION				
Sample ID				
	11/12/24			
	11/13/24 RIRO			
	11/13/24	MM/DD/YY		11/30/26
EXPIRATION DATE				11/30/26
		MM/DD/YYYY		
Sample ID				
	11/13/24			
	RIRO			
	11/13/24			_
NORMALITY USING		0.1000	0.0985 - 0.1015	0.1004
AUTO-TITRATOR		EQUIV/LITER		
Sample ID				
Tested	11/13/24			
Tester	RIRO			
Recorded	11/13/24			
Lot # 411061	115			
11/12	/ 0 4			



Page: 3

Date: 11/14/24 at 7:02 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager