

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

## Customer PO # :

Packaged Product	• DC943F			
Packaged Product	•			
	CYANIDE	STANDARD 0.15 PPM		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	11/12/24			
Tester	MMAT			
Recorded	11/12/24			
CALCULATED PPM CY	ANIDE (CN)	0.15	0.13 - 0.17	0.15
		MG/L		
Sample ID				
Tested	11/12/24			
Tester	MMAT			
Recorded	11/12/24			
APPEARANCE AND CC	LOR	CLEAR WATER WHITE		PASS
		APPEARANCE		
Sample ID				
Tested	11/12/24			
Tester	MMAT			
Recorded	11/12/24			
INSTRUMENTS USED DURING		INSTRUMENT		GC,P4,VF1-Z39
PREPARATION				
Sample ID				
Tested	11/12/24			
Tester	MMAT			
Recorded	11/12/24			
EXPIRATION DATE		MM/DD/YYYY		02/28/2025
		MM/DD/YYYY		
Sample ID				
Tested	11/12/24			
Tester	MMAT			
Recorded	11/12/24			
Lot # 41112	206			
Made 11/12				
made / -				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA