



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: HAC166F			
Hardness Cal Std Ref2 250mg/L CaCO3 EZ5011 Series			
Test	Target/UOM	Range	Result
APPEARANCE AND COLOR			
	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested	11/14/24		
Tester	AROS		
Recorded	11/14/24		
PREPARED TO FORMULATION ON FILE			
	YES YES		YES
Sample ID			
Tested	11/14/24		
Tester	AROS		
Recorded	11/14/24		
INSTRUMENTS USED DURING PREPARATION			
	INSTRUMENT		VF200-T14,.VF100-K7,VF12-3
Sample ID			
Tested	11/14/24		
Tester	AROS		
Recorded	11/14/24		
EXPIRATION DATE			
	MM/DD/YY MM/DD/YYYY		02/28/25
Sample ID			
Tested	11/14/24		
Tester	AROS		
Recorded	11/14/24		
ACT PPM			
	250.0 PPM	248.0 - 252.0	250.2
Sample ID			
Tested	11/14/24		
Tester	AROS		
Recorded	11/14/24		
Lot #	4111434		
Made	11/14/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.