

Page: 1

Date: 11/15/24 at 11:53 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product:	1795			
	CHLORI	NE STANDARD 1000 PPM AS		-
Test	.=	Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
	11/15/24			
	RIRO			
	11/15/24			
PPM CHLORINE (CL2)		1000.0	995.0 - 1005.0	996.9
		PPM		
Sample ID				
Tested	11/15/24			
Tester	RIRO			
Recorded	11/15/24			
APPEARANCE AND COL	OR	CLEAR PALE YELLOW		PASS
		APPEARANCE		
Sample ID				
•	11/15/24			
	RIRO			
	11/15/24			
Actual Density	,	1.00100	0.99969 - 1.00230	1.00108
		G/ML		
Sample ID		- ,		
•	11/15/24			
	RIRO			
	11/15/24			
INSTRUMENTS USED DU		INSTRUMENT		BID-00488,PID-364
PREPARATION	DRING	INSTITUMENT		DID-00400,1 1D-304
Sample ID				
	11/15/24			
	RIRO			
	11/15/24	AMAIDD AGA		24/22/27
EXPIRATION DATE		MM/DD/YY		01/20/25
		MM/DD/YYYY		
Sample ID				
	11/15/24			
	RIRO			
	11/15/24			
SHELF LIFE OF PRODUC	CT	2 MONTHS		2 MONTHS
		TIME		
Sample ID				
•	11/15/24			
	RIRO			
	11/15/24			
Lot # 411115				
"				
Made 11/15/	24			



Page: 2

Date: 11/15/24 at 11:53 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager