

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged	Product:	DC941F			
		CYANIDE	STOCK STANDARD 10 PPM		
Test			Target/UOM	Range	Result
PREPARED	TO FORMUL	ATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	11/12/24			
	Tester	MMAT			
	Recorded	11/12/24			
APPEARAN	CE AND COL	OR	CLEAR WATER WHITE		PASS
			APPEARANCE		
	Sample ID				
	Tested	11/12/24			
	Tester	MMAT			
	Recorded	11/12/24			
INSTRUME	NTS USED DI	JRING	INSTRUMENT		P170,VF500-X30,GC
PREPARAT	ON				
	Sample ID				
	Tested	11/12/24			
	Tester	MMAT			
	Recorded	11/12/24			
Actual PPM			10.00	9.80 - 10.20	10.03
			PPM		
	Sample ID				
	Tested	11/12/24			
	Tester	MMAT			
	Recorded	11/12/24			
EXPIRATIO	N DATE		MM/DD/YYYY		11/12/2024
			MM/DD/YYYY		
	Sample ID				
	Tested	11/12/24			
	Tester	MMAT			
	Recorded	11/12/24			
Lot	# 411123	105			
Mad		/24			
	-				

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland SPA