

Page: 1

Date: 11/15/24 at 10:53 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results
Customer PO #:

Packaged Produ	ict: DOW744F			
	Silver N	Nitrate 0.005 Normal i	n Acetone NIST Traceable	
Test		Target/UOM	Range	Result
PREPARED TO FOR	MULATION ON	YES		YES
FILE		YES		
Sample	ID			
Tested	11/15/24			
Tester	JRIO			
Recorde	ed 11/15/24			
NORMALITY TO 5 DECIMAL PLACES		0.00500	0.00480 - 0.00520	0.00496
		EQUIV/LITER		
Sample	ID			
Tested	11/15/24			
Tester	JRIO			
Recorde	ed 11/15/24			
N.I.S.T. Traceable to SRM 919 NaCl		YES		YES
		YES		
Sample	ID			
Tested	11/15/24			
Tester	JRIO			
Recorde	ed 11/15/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample	ID			
Tested	11/15/24			
Tester	JRIO			
Recorde	ed 11/15/24			
INSTRUMENTS USED DURING		INSTRUMENT		GC,VF5-M5BID-M52
PREPARATION				
Sample	ID			
Tested	11/15/24			
Tester	JRIO			
Recorde	ed 11/15/24			
EXPIRATION DATE		MM/DD/YYYY		11/30/2025
		MM/DD/YYYY		
Sample	ID			
Tested	11/15/24			
Tester	JRIO			
Recorde	ed 11/15/24			
Lot # 411	10710			
_0, ,,	15/24			
IVIAGE ++/	- · = -			



Page: 2

Date: 11/15/24 at 10:53 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager