

Made 11/18/24

Page: 1

Date: 11/18/24 at 9:47 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results
Customer PO #:

Package	ed Product				
		AMMONIA	STANDARD 100 PPM AS NE		
Test			Target/UOM	Range	Result
	ED TO FORMU	JLATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	11/18/24			
	Tester	AROS			
	Recorded	11/18/24			
CERTIFIE	D PPM NH3		100.0	96.0 - 104.0	101.5
			PPM		
	Sample ID				
	Tested	11/18/24			
	Tester	AROS			
	Recorded	11/18/24			
APPEARANCE AND COLOR			CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
	Tested	11/18/24			
	Tester	AROS			
I	Recorded	11/18/24			
INSTRUM	IENTS USED D		INSTRUMENT		M108,GC,VF6-W9,M107
PREPARA					• •
- 	Sample ID				
İ	Tested	11/18/24			
İ	Tester	AROS			
	Recorded	11/18/24			
EXPIRATION DATE		,,	MM/DD/YYYY		02/28/2025
	0		MM/DD/YYYY		-
I	Sample ID		, 2 2,		
I	Tested	11/18/24			
I	Tester	AROS			
I	Recorded	11/18/24			
SHELF LIFE OF PRODUCT			3 MONTHS		3 MONTHS
JIILLI L		,01	TIME		0 1010141110
I	Sample ID		I IIVIL		
I	Tested	11/18/24			
I	Tester	11/18/24 AROS			
I	Recorded	11/18/24			
ļ		538			



Page: 2

Date: 11/18/24 at 9:47 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager