



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: MIL043F			
SILVER NITRATE 0.005 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 11/19/24			
Tester JRIO			
Recorded 11/19/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID	YES		
Tested 11/19/24			
Tester JRIO			
Recorded 11/19/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		11/19/2024
Sample ID	MM/DD/YYYY		
Tested 11/19/24			
Tester JRIO			
Recorded 11/19/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	20.0
Sample ID	DEG C		
Tested 11/19/24			
Tester JRIO			
Recorded 11/19/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		11/19/2024
Sample ID	MM/DD/YYYY		
Tested 11/19/24			
Tester JRIO			
Recorded 11/19/24			
AMBIENT TEST TEMPERATURE DEG C	20.0	15.0 - 30.0	20.0
Sample ID	DEG C		
Tested 11/19/24			
Tester JRIO			
Recorded 11/19/24			
Normality Reported to 5 decimal	0.00500	0.00480 - 0.00520	0.00497
Sample ID	EQ/L		
Tested 11/19/24			
Tester JRIO			
Recorded 11/19/24			
QCSP54:STANDARDIZATION OF LOW NORMAL SN	YES		YES
Sample ID	QCSP54		
Tested 11/19/24			
Tester JRIO			
Recorded 11/19/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 919 NaCl	YES YES	YES
Sample ID		
Tested	11/19/24	
Tester	JRIO	
Recorded	11/19/24	
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	11/19/24	
Tester	JRIO	
Recorded	11/19/24	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M112,GC,T-223,BID-M92,PID-3 40
Sample ID		
Tested	11/19/24	
Tester	JRIO	
Recorded	11/19/24	
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	11/30/2025
Sample ID		
Tested	11/19/24	
Tester	JRIO	
Recorded	11/19/24	
Lot #	4110814	
Made	11/15/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager