



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8407			
SILVER NITRATE 0.025 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0250 EQ/L	0.0245 - 0.0255	0.0250
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>QCSP53: SILVER NITRATE STD ON AUTO-TITRA</b>	YES METHOD		YES
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>QCSP54: STANDARDIZATION OF LOW NORMAL SN</b>	YES QCSP54		YES
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>QCSP78: CHLORIDE SQC METHOD TO VERIFY SN</b>	YES QCSP78		YES
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>N.I.S.T. Traceable to SRM 919</b>	Yes YES		YES
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>PREPARED TO FORMULATION ON FILE</b>	YES YES		YES
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES YES		YES
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY MM/DD/YYYY		11/18/2024
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0 DEG C	10.0 - 35.0	20.4
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY MM/DD/YYYY		11/18/2024
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0 DEG C	15.0 - 30.0	20.8
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>Actual Density</b>	1.00180 G/ML	1.00140 - 1.00219	1.00182
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M55,M112,T157,M91,M52,TID5 3309
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		
<b>EXPIRATION DATE</b>	MM/DD/YYYY MM/DD/YYYY		05/30/2026
Sample ID			
Tested	11/18/24		
Tester	LWIL		
Recorded	11/18/24		

**Lot #** 4110813

**Made** 11/15/24



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536  
Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 3

Date: 11/18/24 at 1:59 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**