



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: SPE949F			
POTASSIUM HYDROXIDE 4.65 ± 0.015 NORMAL (1L)			
Test	Target/UOM	Range	Result
FORMULATION INCLUDES ULTRA PURE WATER	YES		YES
Sample ID			
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		11/22/2024
Sample ID			
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	19.1
Sample ID			
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		11/22/2024
Sample ID			
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	19.1
Sample ID			
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M113,T-169,BID-00488,PID-M9
Sample ID			1
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

NORMALITY TO 4 DEC PLACES	4.6500	4.6350 - 4.6650	4.6472
	EQUIV/LITER		
Sample ID			
Tested	11/25/24		
Tester	RIRO		
Recorded	11/25/24		
Actual Density	1.20318	1.20170 - 1.20465	1.20349
	G/ML		
Sample ID			
Tested	11/25/24		
Tester	RIRO		
Recorded	11/25/24		
N.I.S.T. Traceable to SRM 84	YES		YES
	YES		
Sample ID			
Tested	11/25/24		
Tester	RIRO		
Recorded	11/25/24		
ULTRA PURE WATER USED IN PREP	YES		YES
	YES		
Sample ID			
Tested	11/25/24		
Tester	RIRO		
Recorded	11/25/24		
SHELF LIFE OF PRODUCT	2 YEARS		2 YEARS
	TIME		
Sample ID			
Tested	11/25/24		
Tester	RIRO		
Recorded	11/25/24		
Lot #	4111505		
Made	11/25/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager