

Page: 1

Date: 11/26/24 at 2:42 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc	t: 1540			
_	SODIUM 1	REFERENCE BUFFER PH 2.0	00 ± 0.02	
Test		Target/UOM	Range	Result
PREPARED TO FORM	IULATION ON	YES		YES
FILE		YES		
Sample II	D			
Tested	11/26/24			
Tester	DTRA			
Recorded				
PH OF PRODUCT/SOLUTION		2.000	1.980 - 2.020	2.002
		PH		
Sample II				
Tested	11/26/24			
Tester	DTRA			
Recorded	I 11/26/24			
Actual Temp		25.0	24.0 - 26.0	25.0
		DEG C		
Sample II				
Tested	11/26/24			
Tester	DTRA			
Recorded				
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample II				
Tested	11/26/24			
Tester	DTRA			
Recorded				
INSTRUMENTS USED	DURING	INSTRUMENT		M103,T28A
PREPARATION				
Sample II				
Tested	11/26/24			
Tester	DTRA			
Recorded	I 11/26/24	V50		\/F0
N.I.S.T. Traceable		YES		YES
	_	YES		
Sample II				
Tested	11/26/24			
Tester	DTRA			
Recorded EXPIRATION DATE	I 11/26/24	MM/DD/VVV		11/30/2026
Sample ID		MM/DD/YYYY		11/30/2020
-				
Tested	11/26/24			
Tester	DTRA I 11/26/24			
Recorded				
Lot # 4112				
Made 11/2	26/24			



Page: 2

Date: 11/26/24 at 2:42 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager