



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: DC839F			
TOC/TN STANDARD 25 PPM EACH COMPONENT (MAKE FRESH)			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 11/22/24			
Tester DTRA			
Recorded 11/22/24			
CERTIFIED PPM TOC	25.0 PPM	23.0 - 27.0	25.3
Sample ID			
Tested 11/22/24			
Tester DTRA			
Recorded 11/22/24			
CALCULATED PPM NITROGEN	25.0 PPM	24.0 - 26.0	25.0
Sample ID			
Tested 11/22/24			
Tester DTRA			
Recorded 11/22/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 11/22/24			
Tester DTRA			
Recorded 11/22/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P174,P101VF12-3
Sample ID			
Tested 11/22/24			
Tester DTRA			
Recorded 11/22/24			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		01/27/2025
Sample ID			
Tested 11/22/24			
Tester DTRA			
Recorded 11/22/24			
Lot #	4112163		
Made	11/22/24		



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 11/22/24 at 1:05 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager