

Page: 1

Date: 11/22/24 at 2:23 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: 0371			
	Ammonia	Nitrogen Standard 10 F	PPM as NH3	
Test		Target/UOM	Range	Result
PREPARED TO FORMULATION ON		YES	<del>-</del>	YES
FILE		YES		
Sample ID				
Tested	11/22/24			
Tester	JTRA			
Recorded	11/22/24			
CERTIFIED PPM NH3		10.0	9.0 - 11.0	9.9
		PPM		
Sample ID				
Tested	11/22/24			
Tester	JTRA			
Recorded	11/22/24			
APPEARANCE AND CO		CLEAR LIQUID		PASS
74 1 2744 4162 7415 662611		APPEARANCE		17.00
Sample ID		ALLEAVANOL		
Tested	11/22/24			
	JTRA			
Tester				
Recorded	11/22/24	INICTOLIMENT		D290 CC VEC WE
INSTRUMENTS USED DURING		INSTRUMENT		P389,GC,VF6-W5
PREPARATION				
Sample ID				
Tested	11/22/24			
Tester	JTRA			
Recorded	11/22/24			
EXPIRATION DATE		MM/DD/YY		02/28/25
		MM/DD/YYYY		
Sample ID				
Tested	11/22/24			
Tester	JTRA			
Recorded	11/22/24			
SHELF LIFE OF PRODUCT		SHELF LIFE		3 MONTHS
		TIME		
Sample ID				
Tested	11/22/24			
Tester	JTRA			
Recorded	11/22/24			
Lot # 41122				
Made 11/22				
IVIAUE ±±/22	_, _ 1			



Page: 2

Date: 11/22/24 at 2:23 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**