Page: 1

Date: 11/22/24 at 2:08 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packag	ed Product:				
		NITROG	EN STD 100 PPM (1ML = 0 .		
Test			Target/UOM	Range	Result
	ED TO FORMU	LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	11/22/24			
	Tester	JTRA			
	Recorded	11/22/24			
CERTIFIED PPM NITROGEN:		GEN:	100.0	96.0 - 104.0	98.7
			PPM		
İ	Sample ID				
	Tested	11/22/24			
	Tester	JTRA			
	Recorded	11/22/24			
APPEARANCE AND COLOR		LOR	CLEAR WATER WHITE		PASS
			LIQUID		
			APPEARANCE		
	Sample ID				
	Tested	11/22/24			
	Tester	JTRA			
	Recorded	11/22/24			
INSTRUM	ENTS USED D	URING	INSTRUMENT		M115,GC,VF4-N21
PREPARA	ATION				
	Sample ID				
	Tested	11/22/24			
	Tester	JTRA			
	Recorded	11/22/24			
FXPIRAT	ION DATE	,,	MM/DD/YYYY		02/28/2025
_XI IIIXII			MM/DD/YYYY		02, 20, 2020
	Sample ID				
	Tested	11/22/24			
	Tester	11/22/24 JTRA			
	Recorded	11/22/24			
	ot# 41121				
Ma	ade $11/22$:/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1.