

Page: 1

Date: 11/18/24 at 5:22 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product: 8766			
	SULFITE 1.0 MOLAR pH 9	.900 ± 0.100	
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 11/18/24			
Tester DNEW			
Recorded 11/18/24			
CALCULATED MOLARITY	1.000	0.992 - 1.008	1.000
	MOL/L		
Sample ID			
Tested 11/18/24			
Tester DNEW			
Recorded 11/18/24			
APPEARANCE AND COLOR	PALE YELLOW		PASS
	APPEARANCE		
Sample ID			
Tested 11/18/24			
Tester DNEW			
Recorded 11/18/24			
INSTRUMENTS USED DURING	INSTRUMENT		M112,T-4
PREPARATION			
Sample ID			
Tested 11/18/24			
Tester DNEW			
Recorded 11/18/24			
PH OF PRODUCT/SOLUTION	9.900	9.800 - 10.000	9.881
	PH		
Sample ID			
Tested 11/18/24			
Tester DNEW			
Recorded 11/18/24			
EXPIRATION DATE	MM/DD/YY		05/30/25
	MM/DD/YYYY		
Sample ID			
Tested 11/18/24			
Tester DNEW			
Recorded 11/18/24			
SHELF LIFE OF PRODUCT	6 MONTHS		6 MONTHS
	TIME		
Sample ID			
Tested 11/18/24			
Tester DNEW			
Recorded 11/18/24			
Lot # 4111445			
Made 11/18/24			



Page: 2

Date: 11/18/24 at 5:22 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager