



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

Test	Target/UOM	Range	Result
<b>Packaged Product:</b> DC917F			
	SODIUM HYDROXIDE 0.204 NORMAL AQUEOUS CERTIFIED		
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 11/20/24			
Tester JRIO			
Recorded 11/20/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 11/20/24			
Tester JRIO			
Recorded 11/20/24			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		11/20/2024
Sample ID			
Tested 11/20/24			
Tester JRIO			
Recorded 11/20/24			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	20.0
Sample ID			
Tested 11/20/24			
Tester JRIO			
Recorded 11/20/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		11/20/2024
Sample ID			
Tested 11/20/24			
Tester JRIO			
Recorded 11/20/24			
<b>AMBIENT TEST TEMPERATURE DEG C</b>	20.0	15.0 - 30.0	20.0
Sample ID			
Tested 11/20/24			
Tester JRIO			
Recorded 11/20/24			
<b>NORMALITY TO 4 DEC PLACES</b>	0.2040	0.2035 - 0.2045	0.2043
EQUIV/LITER			
Sample ID			
Tested 11/20/24			
Tester JRIO			
Recorded 11/20/24			
<b>N.I.S.T. Traceable to SRM 84</b>	YES		YES
Sample ID			
Tested 11/20/24			
Tester JRIO			
Recorded 11/20/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested 11/20/24		
Tester JRIO		
Recorded 11/20/24		
<b>Actual Density</b>	1.00741 G/ML	1.00724 - 1.00758
Sample ID		1.00747
Tested 11/20/24		
Tester JRIO		
Recorded 11/20/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT	M55,M112,T172
Sample ID		
Tested 11/20/24		
Tester JRIO		
Recorded 11/20/24		
<b>EXPIRATION DATE</b>	MM/DD/YYYY	11/30/2026
Sample ID	MM/DD/YYYY	
Tested 11/20/24		
Tester JRIO		
Recorded 11/20/24		
<b>Lot #</b>	4111216	
<b>Made</b>	11/15/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**