

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product:	6921			
	CONDUCT	SIVITY STANDARD 10.0 uMHO	S	
Test		Target/UOM	Range	Result
PREPARED TO FORMUL	ATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	11/21/24			
Tester	RIRO			
Recorded	11/21/24			
CONDUCTIVITY @ 25°C		10.0	9.9 - 10.1	10.0
PLACE		US/CM		
Sample ID				
Tested	11/21/24			
Tester	RIRO			
Recorded	11/21/24			
QCSP30:CONDUCTIVITY		YES		YES
STD&MEASUREMENT		QCSP30		
Sample ID				
Tested	11/21/24			
Tester	RIRO			
	11/21/24			
TRACEABLE TO N.I.S.T.		YES		YES
		YES		
Sample ID				
Tested	11/21/24			
Tester	RIRO			
	11/21/24			
APPEARANCE AND COL		CLEAR WATER WHITE		PASS
/ 1 2/ 40/4102 / 412 002		LIQUID		
		APPEARANCE		
Sample ID				
Tested	11/21/24			
Tester	RIRO			
	11/21/24			
INSTRUMENTS USED D		INSTRUMENT		T229
PREPARATION				1220
Sample ID				
Tested	11/21/24			
Tester	RIRO			
Recorded	11/21/24			
EXPIRATION DATE	11/21/24	MM/DD/YY		02/28/25
		MM/DD/YYYY		02/20/20
Sample ID				
Tested	11/21/24			
Tester	RIRO			
	11/21/24			
SHELF LIFE OF PRODU		SHELF LIFE		3 MONTHS
	.	TIME		
Sample ID				
Tested	11/21/24			
Tester	RIRO			
Recorded	11/21/24			



Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Lot #	4112061
Made	11/21/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

chard Sto

Quality Manager