

Page: 1

Date: 11/22/24 at 1:57 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Produc				
	CONDUC'	TIVITY STANDARD 4,600 uM		
Test		Target/UOM	Range	Result
PREPARED TO FORM	IULATION ON	YES		YES
FILE		YES		
Sample II)			
Tested	11/22/24			
Tester	RIRO			
Recorded				
CONDUCTIVITY @25	C.	4600	4599 - 4601	4600
		US/CM		
Sample II)			
Tested	11/22/24			
Tester	AROS			
Recorded	11/22/24			
QCSP30:CONDUCTIV	ITY	YES		YES
STD&MEASUREMENT	Г	QCSP30		
Sample II)			
Tested	11/22/24			
Tester	AROS			
Recorded				
TRACEABLE TO N.I.S		YES		YES
		YES		
Sample II)			
Tested	11/22/24			
Tester	AROS			
Recorded				
APPEARANCE AND C		CLEAR WATER WHITE		PASS
A I LAIVANGE AND G	OLOK	LIQUID		17.65
		APPEARANCE		
Camania II	_	AFFLANANCE		
Sample II				
Tested	11/22/24			
Tester	AROS			
Recorded		INICEDIMENT		MOO TOOO
INSTRUMENTS USED	DURING	INSTRUMENT		M92,T229
PREPARATION	_			
Sample II				
Tested	11/22/24			
Tester	AROS			
Recorded	11/22/24			
EXPIRATION DATE		MM/DD/YYYY		11/30/2025
		MM/DD/YYYY		
Sample II				
Tested	11/22/24			
Tester	AROS			
Recorded				
Lot # 4112	2139			
Made 11/2	22/24			



Page: 2

Date: 11/22/24 at 1:57 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager