



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: HA-20760-32			
MOLYBOVANADATE REAGENT(100mls) FOR PHOSPHATES			
Test	Target/UOM	Range	Result
Function Test	0.1110	0.1000 - 0.1220	0.1059
	ABSORBANCE		
Sample ID			
Tested	12/03/24		
Tester	TBUR		
Recorded	12/03/24		
Function test concentration blank	0.483-0.549 ABS		0.5199
	ABS		
Sample ID			
Tested	12/03/24		
Tester	TBUR		
Recorded	12/03/24		
Reagent blank	0.0185	0.0000 - 0.0370	0.0024
	ABSORBANCE(ABS)		
Sample ID			
Tested	12/03/24		
Tester	TBUR		
Recorded	12/03/24		
EXPIRATION DATE	MM/DD/YYYY		08/30/2029
	MM/DD/YYYY		
Sample ID			
Tested	12/03/24		
Tester	TBUR		
Recorded	12/03/24		
Lot #	433110		
Made	11/26/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Richard [Signature]

Quality Manager