



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: HA-22538-69			
AMINO ACID F POWDER PILLOWS FOR 25ML SAMPLE ((PK/100))			
Test	Target/UOM	Range	Result
Absorbance of a 3 ppm silica std v blank	1.022 - 1.168 ABS		1.1080
Sample ID			
Tested 12/03/24			
Tester TBUR			
Recorded 12/03/24			
Absorbance of a 1 ppm silica std v blank	0.336-0.394 ABS		0.3770
Sample ID			
Tested 12/03/24			
Tester TBUR			
Recorded 12/03/24			
Absorbance of standard vs the reagent bl	- 0.012 - 0.012 ABS		0.0030
Sample ID			
Tested 12/03/24			
Tester TBUR			
Recorded 12/03/24			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		10/30/2027
Sample ID			
Tested 12/03/24			
Tester TBUR			
Recorded 12/03/24			
Lot #	433117		
Made	11/26/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager