

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Product	M-236F			
3		C MOBILE PHASE		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	ILATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	12/13/24			
Tester	DNEW			
Recorded	12/13/24			
PH OF PRODUCT/SOLUTION		2.700	2.680 - 2.720	2.699
		PH		
Sample ID				
Tested	12/13/24			
Tester	DNEW			
Recorded	12/13/24			
FILTERED THROUGH	0.2 MICRON	YES		YES
FILTER		YES		
Sample ID				
Tested	12/13/24			
Tester	DNEW			
Recorded	12/13/24			
APPEARANCE AND CO	LOR	LIGHT BLUE LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	12/13/24			
Tester	DNEW			
Recorded	12/13/24			
Actual Density		0.99895	0.99690 - 1.00099	0.99860
-		G/ML		
Sample ID				
Tested	12/13/24			
Tester	DNEW			
Recorded	12/13/24			
INSTRUMENTS USED	DURING	INSTRUMENT		M108,M102,T-14,DP
PREPARATION				
Sample ID				
Tested	12/13/24			
Tester	DNEW			
Recorded	12/13/24			
EXPIRATION DATE		MM/DD/YYYY		06/30/2025
		MM/DD/YYYY		
Sample ID				
Tested	12/13/24			
Tester	DNEW			
Recorded	12/13/24			
Lot # 41127	777			
Made 12/13				
	~,			



Page: 2 Date: 12/13/24 at 2:45 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results Customer PO # :

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1 chard SE

Quality Manager