



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 3550			
EDTA TITRANT 0.01 MOLAR (0.02 NORMAL) (1ML = 1MG CaCO3) NIST			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 12/06/24			
Tester JRIO			
Recorded 12/06/24			
QCSP4:STANDARDIZATION EDTA SOLUTIONS	YES		YES
Sample ID	QCSP4		
Tested 12/06/24			
Tester JRIO			
Recorded 12/06/24			
Actual Density	1.00009	0.99990 - 1.00028	1.00018
	G/ML		
Sample ID			
Tested 12/06/24			
Tester JRIO			
Recorded 12/06/24			
N.I.S.T. Traceable to SRM 915	YES		YES
	YES		
Sample ID			
Tested 12/06/24			
Tester JRIO			
Recorded 12/06/24			
Prepared to API 13A Sec. 7.11.2	YES		YES
	YES		
Sample ID			
Tested 12/06/24			
Tester JRIO			
Recorded 12/06/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 12/06/24			
Tester JRIO			
Recorded 12/06/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		T209,M112,M55
Sample ID			
Tested 12/06/24			
Tester JRIO			
Recorded 12/06/24			
EXPIRATION DATE	MM/DD/YY		12/30/26
	MM/DD/YYYY		
Sample ID			
Tested 12/06/24			
Tester JRIO			
Recorded 12/06/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Molarity	0.0100 MOL/L	0.0095 - 0.0105	0.0104
Sample ID			
Tested	12/06/24		
Tester	JRIO		
Recorded	12/06/24		
SHELF LIFE OF PRODUCT	24 MONTHS TIME		24 MONTHS
Sample ID			
Tested	12/06/24		
Tester	JRIO		
Recorded	12/06/24		
Lot #	4112029		
Made	12/06/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager