



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7257			
POTASSIUM HYDROXIDE 0.05 NORMAL IN IPA			
Test	Target/UOM	Range	Result
NORMALITY TO 4 DEC PLACES	0.0500 EQUIV/LITER	0.0495 - 0.0505	0.0501
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
SAMPLE COLLECTION & HANDLING PER LAB040	YES YES		YES
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY MM/DD/YYYY		12/05/2025
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	19.9
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		12/05/2025
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
AMBIENT TEST TEMPERATURE DEG C	20.0 DEG C	15.0 - 30.0	19.9
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
QCSP7: STANDARDIZATION OF AQUEOUS BASES	YES METHOD		YES
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

N.I.S.T. Traceable to SRM 84 (KHP)	YES		YES
	YES		
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE TO LIGHT HAZE APPEARANCE		PASS
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M120,T16
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
Actual Density	0.78926 G/ML	0.78859 - 0.78994	0.78925
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/25
Sample ID			
Tested	12/05/24		
Tester	LWIL		
Recorded	12/05/24		
Lot #	41115100		
Made	12/05/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager