Page: 1

Date: 12/12/24 at 4:26 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Pro	oduct:	EQS301F			
		TOC STANDAR	RD CHECK 150 PPM		
Test			Target/UOM	Range	Result
PREPARED TO F	FORMULA [*]		YES		YES
FILE			YES		
Sam	nple ID				
Test	ted 12	2/12/24			
Test	ter D	TRA			
Rec	orded 12	2/12/24			
ACTUAL PPM CERTIFIED TOC		тос	150	146 - 154	150
VALUE			PPM		
Sam	nple ID				
Test	-	2/12/24			
Test		TRA			
Rec	orded 12	2/12/24			
APPEARANCE AND COLOR			CLEAR LIQUID		PASS
			APPEARANCE		
Sam	nple ID				
Test	-	2/12/24			
Test		TRA			
Rec		2/12/24			
INSTRUMENTS USED DURING		RING	LAB EQUIPMENT		M-81,VF5-L15
PREPARATION			INSTRUMENT		
Sam	nple ID				
Test	-	2/12/24			
Test		TRA			
Rec	orded 12	2/12/24			
EXPIRATION DATE			MM/DD/YYYY		06/30/2025
			MM/DD/YYYY		
Sam	nple ID				
Test	-	2/12/24			
Test		TRA			
		2/12/24			
Lot # 4120937					
	12/12/2	24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Prichard St.