Page: 1

Date: 12/02/24 at 7:42 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Package	d Product:	SPE800F			
	COD STANDARD 150 PPM				
Test			Target/UOM	Range	Result
PREPARE	D TO FORMULA	ATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested 1	12/02/24			
	Tester [OTRA			
	Recorded 1	12/02/24			
PPM COD			150	137 - 162	147
			PPM		
	Sample ID				
		12/02/24			
	Tester [OTRA			
	Recorded 1	12/02/24			
APPEARANCE AND COLOR		OR .	CLEAR LIQUID		PASS
			APPEARANCE		
	Sample ID				
		12/02/24			
		OTRA			
		12/02/24			
INSTRUMENTS USED DURING		RING	INSTRUMENT		P101,P173,VF1-Z32
PREPARA	TION				
	Sample ID				
		12/02/24			
	Tester [OTRA			
		12/02/24			
EXPIRATION	ON DATE		MM/DD/YYYY		02/28/2025
			MM/DD/YYYY		
	Sample ID				
		12/02/24			
		OTRA			
		12/02/24			
	t# 411224				
Ma	ae TZ/UZ/	∠ 4			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

huland I