Page: 1

Date: 12/10/24 at 6:11 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged	Product:	LY162F			
			UM STANDARD 1.0 PPM IN 1	METHANOL	
Test			Target/UOM	Range	Result
PREPARED	TO FORMU	LATION ON	YES		YES
FILE			YES		
	Sample ID				
	Tested	12/10/24			
	Tester	AROS			
	Recorded	12/10/24			
РРМ К			1.00	0.98 - 1.02	1.00
			PPM		
	Sample ID				
	Tested	12/10/24			
	Tester	AROS			
	Recorded	12/10/24			
APPEARANCE AND COLOR		_OR	CLEAR WATER WHITE		PASS
			APPEARANCE		
	Sample ID				
	Tested	12/10/24			
	Tester	AROS			
	Recorded	12/10/24			
INSTRUMEN			LAB EQUIPMENT		P403,VF500-X31
PREPARATION			INSTRUMENT		,
	Sample ID				
	Tested	12/10/24			
	Tester	AROS			
	Recorded	12/10/24			
EXPIRATION DATE			MM/DD/YYYY		01/15/2025
			MM/DD/YYYY		0 // 10/2020
	Sample ID		, 2 2,		
	Tested	12/10/24			
	Tester	AROS			
	Recorded	12/10/24			
l of	# 41210				
Mad	e 12/10	/ 4			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Parland St