

Recorded 12/12/24

Page: 1

Date: 12/12/24 at 6:01 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:				
Packaged Product:	1764			
-	CHLORIDE	STANDARD 1000 PPM a	s Cl Certified	
Test		Target/UOM	Range	Result
PREPARED TO FORMU	LATION ON	YES		YES
ILE		YES		
Sample ID				
Tested	12/12/24			
Tester	JTRA			
Recorded	12/12/24			
CHLORIDE PPM		1000.0	980.0 - 1020.0	1006.3
		PPM		
Sample ID				
Tested	12/12/24			
Tester	JTRA			
Recorded	12/12/24			
N.I.S.T. Traceable to SRM 919 NaCl		YES		YES
		YES		
Sample ID				
Tested	12/12/24			
Tester	JTRA			
Recorded	12/12/24			
APPEARANCE AND CO	LOR	CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	12/12/24			
Tester	JTRA			
Recorded	12/12/24			
Actual Density		0.99941	0.99930 - 0.99953	0.99941
		G/ML		
Sample ID				
Tested	12/12/24			
Tester	JTRA			
Recorded	12/12/24			
INSTRUMENTS USED DURING		INSTRUMENT		M115,VF12-103
PREPARATION				
Sample ID				
Tested	12/12/24			
Tester	JTRA			
Recorded	12/12/24			
EXPIRATION DATE		MM/DD/YY		12/30/25
		MM/DD/YYYY		
Sample ID				
Tested	12/12/24			
Tester	JTRA			
Recorded	12/12/24			
SHELF LIFE OF PRODU	ICT	1 YEAR		1 YEAR
		TIME		
Sample ID				
Tested	12/12/24			
Tester	JTRA			
B	10/10/01			

Page: 2

Date: 12/12/24 at 6:01 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Lot # 4112120 Made 12/12/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager