



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8667			
SODIUM HYDROXIDE 1.2 NORMAL LOW CARBONATE			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID			
Tested 12/03/24			
Tester DTRA			
Recorded 12/03/24			
<b>SAMPLE COLLECTION &amp; HANDLING PER LAB040</b>	YES		YES
Sample ID			
Tested 12/03/24			
Tester DTRA			
Recorded 12/03/24			
<b>DATE OF SAMPLING FOR TEST PURPOSES</b>	MM/DD/YYYY		12/03/2024
Sample ID			
Tested 12/03/24			
Tester DTRA			
Recorded 12/03/24			
<b>AMBIENT TEMPERATURE DURING SAMPLING DEGC</b>	25.0	10.0 - 35.0	19.0
Sample ID			
Tested 12/03/24			
Tester DTRA			
Recorded 12/03/24			
<b>DATE OF TESTING OF SAMPLE</b>	MM/DD/YYYY		12/03/2024
Sample ID			
Tested 12/03/24			
Tester DTRA			
Recorded 12/03/24			
<b>AMBIENT TEST TEMPERATURE DEGC</b>	20.0	15.0 - 30.0	19.0
Sample ID			
Tested 12/03/24			
Tester DTRA			
Recorded 12/03/24			
<b>Product nitrogen purged</b>	YES		YES
Sample ID			
Tested 12/03/24			
Tester DTRA			
Recorded 12/03/24			
<b>QCSP7: STANDARDIZATION OF AQUEOUS BASES</b>	YES		YES
Sample ID			
Tested 12/03/24			
Tester DTRA			
Recorded 12/03/24			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>Actual Normality</b>	1.2000	1.1990 - 1.2010	1.2000
	EQ/L		
Sample ID			
Tested	12/03/24		
Tester	DTRA		
Recorded	12/03/24		
<b>N.I.S.T. Traceable to SRM 84 (KHP)</b>	YES		YES
	YES		
Sample ID			
Tested	12/03/24		
Tester	DTRA		
Recorded	12/03/24		
<b>PPM TOC TO 2 DEC PLACES</b>	5.00	0.00 - 5.00	0.19
	PPM		
Sample ID			
Tested	12/03/24		
Tester	DTRA		
Recorded	12/03/24		
<b>PPM INORGANIC CARBON</b>	25.0	0.0 - 25.0	0.8
	PPM		
Sample ID			
Tested	12/03/24		
Tester	DTRA		
Recorded	12/03/24		
<b>Actual Density</b>	1.04902	1.04852 - 1.04952	1.04913
	G/ML		
Sample ID			
Tested	12/03/24		
Tester	DTRA		
Recorded	12/03/24		
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested	12/03/24		
Tester	DTRA		
Recorded	12/03/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		T8667,M55
Sample ID			
Tested	12/03/24		
Tester	DTRA		
Recorded	12/03/24		
<b>SHELF LIFE OF PRODUCT</b>	1 YEAR		1 YEAR
	TIME		
Sample ID			
Tested	12/03/24		
Tester	DTRA		
Recorded	12/03/24		



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer PO # :

<b>EXPIRATION DATE</b>	MM/DD/YYYY	12/30/2025
	MM/DD/YYYY	
Sample ID		
Tested	12/03/24	
Tester	DTRA	
Recorded	12/03/24	
<b>Lot #</b>	4112610	
<b>Made</b>	12/03/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**