



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 3540			
EDTA TITRANT 0.1 MOLAR (0.2N)			
Test	Target/UOM	Range	Result
<b>Actual Molarity</b>	0.1000 MOL/L	0.0995 - 0.1005	0.0996
Sample ID			
Tested	12/09/24		
Tester	LWIL		
Recorded	12/09/24		
<b>QCSP4:STANDARDIZATION EDTA SOLUTIONS</b>	YES QCSP4		YES
Sample ID			
Tested	12/09/24		
Tester	LWIL		
Recorded	12/09/24		
<b>PREPARED TO FORMULATION ON FILE</b>	YES YES		YES
Sample ID			
Tested	12/09/24		
Tester	LWIL		
Recorded	12/09/24		
<b>FORMULA WEIGHT</b>	372.24 F.W.	= 372.24	372.24
Sample ID			
Tested	12/09/24		
Tester	LWIL		
Recorded	12/09/24		
<b>APPEARANCE AND COLOR</b>	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested	12/09/24		
Tester	LWIL		
Recorded	12/09/24		
<b>Actual Density</b>	1.01660 G/ML	1.01584 - 1.01736	1.01677
Sample ID			
Tested	12/09/24		
Tester	LWIL		
Recorded	12/09/24		
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M87,T-9,BID-00787,PID-M91
Sample ID			
Tested	12/09/24		
Tester	LWIL		
Recorded	12/09/24		
<b>Traceable to N.I.S.T. SRM 915</b>	YES YES		YES
Sample ID			
Tested	12/09/24		
Tester	LWIL		
Recorded	12/09/24		




**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>EXPIRATION DATE</b>	MM/DD/YY MM/DD/YYYY	12/30/26
Sample ID		
Tested	12/09/24	
Tester	LWIL	
Recorded	12/09/24	
<b>SHELF LIFE OF PRODUCT</b>	<b>2 YEARS</b> TIME	<b>2 YEARS</b>
Sample ID		
Tested	12/09/24	
Tester	LWIL	
Recorded	12/09/24	
<b>Lot #</b>	4112524	
<b>Made</b>	12/09/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



**Quality Manager**