



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: CY394 Conductivity Blind QA Standard (1L)			
Lot # 4111835			
Chemical Product: CY394F Conductivity Blind QA Standard			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE			
	YES		YES
	YES		
Sample ID			
Tested 11/22/24			
Tester RIRO			
Recorded 11/22/24			
CONDUCTIVITY @ 25°C 1 DEC PLACE			
	17.5	10.0 - 25.0	11.7
	US/CM		
Sample ID			
Tested 12/03/24			
Tester CCOA			
Recorded 12/03/24			
APPEARANCE AND COLOR			
	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			
INSTRUMENTS USED DURING PREPARATION			
	INSTRUMENT		T-234,M92
Sample ID			
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			
EXPIRATION DATE			
	MM/DD/YYYY		02/28/2025
	MM/DD/YYYY		
Sample ID			
Tested 11/25/24			
Tester RIRO			
Recorded 11/25/24			
Made 11/25/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 2

Date: 12/03/24 at 9:49 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager