



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 7330			
POTASSIUM IODIDE-IODATE 0.025 NORMAL			
Test	Target/UOM	Range	Result
Actual Normality	0.0250 EQ/L	0.0245 - 0.0255	0.0249
Sample ID			
Tested	12/02/24		
Tester	JENC		
Recorded	12/02/24		
QCSP33:POTASSIUM IODATE-IODIDE TITRANT	YES QCSP33		YES
Sample ID			
Tested	12/02/24		
Tester	JENC		
Recorded	12/02/24		
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested	12/02/24		
Tester	JENC		
Recorded	12/02/24		
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
Sample ID			
Tested	12/02/24		
Tester	JENC		
Recorded	12/02/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M91,M112,T230
Sample ID			
Tested	12/02/24		
Tester	JENC		
Recorded	12/02/24		
Actual Density	1.00577 G/ML	1.00558 - 1.00596	1.00577
Sample ID			
Tested	12/02/24		
Tester	JENC		
Recorded	12/02/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/25
Sample ID			
Tested	12/02/24		
Tester	JENC		
Recorded	12/02/24		
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		1 YEAR
Sample ID			
Tested	12/02/24		
Tester	JENC		
Recorded	12/02/24		



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 136

YES

YES

YES

Sample ID

Tested 12/02/24

Tester JENC

Recorded 12/02/24

Lot # 4111413

Made 12/02/24

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager