

Page: 1

Date: 12/10/24 at 3:59 PM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569

Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	• OXY3603E	7		
-		TE SOLUTION		
Test		Target/UOM	Range	Result
PREPARED TO FORMU	JLATION ON	YES		YES
FILE		YES		
Sample ID				
Tested	12/10/24			
Tester	MMAT			
Recorded	12/10/24			
CALCULATED W/V% S	ODIUM	10.0	9.9 - 10.1	10.0
PHOSPHATE MONOBAS		W/V%		
Sample ID				
Tested	12/10/24			
Tester	MMAT			
Recorded	12/10/24			
APPEARANCE AND CO		CLEAR LIQUID		PASS
7 7		APPEARANCE		. ,
Sample ID		711 2711 1110		
Tested	12/10/24			
Tester	MMAT			
Recorded	12/10/24			
INSTRUMENTS USED I		INSTRUMENT		M120,VF2-B17
PREPARATION	JOHING	INSTITUTION		W1120, V1 2-D17
Sample ID	10/10/01			
Tested	12/10/24			
Tester	MMAT			
Recorded	12/10/24	4 00070	1.05000 1.00010	4.0004
Actual Density		1.06070	1.05800 - 1.06340	1.06224
		G/ML		
Sample ID				
Tested	12/10/24			
Tester	MMAT			
Recorded	12/10/24			
EXPIRATION DATE		MM/DD/YYYY		12/30/2025
		MM/DD/YYYY		
Sample ID				
Tested	12/10/24			
Tester	MMAT			
Recorded	12/10/24			
Lot # 41210	032			
Made 12/10				
wade ±2/±0	J / 4 I			



Page: 2

Date: 12/10/24 at 3:59 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO #:

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager