



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8668			
SODIUM HYDROXIDE 8.0 NORMAL CERTIFIED			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 12/12/24			
Tester ENOR			
Recorded 12/12/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 12/12/24			
Tester ENOR			
Recorded 12/12/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	20.6
Sample ID			
Tested 12/12/24			
Tester ENOR			
Recorded 12/12/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		12/12/2024
Sample ID			
Tested 12/12/24			
Tester ENOR			
Recorded 12/12/24			
AMBIENT TEST TEMPERATURE DEGC	20.0 DEG C	15.0 - 30.0	20.0
Sample ID			
Tested 12/12/24			
Tester ENOR			
Recorded 12/12/24			
APPEARANCE AND COLOR	CLEAR WATER WHITE APPEARANCE		PASS
Sample ID			
Tested 12/12/24			
Tester ENOR			
Recorded 12/12/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		VF12-101,BID-08446,PID-176
Sample ID			
Tested 12/12/24			
Tester ENOR			
Recorded 12/12/24			
Actual Normality	8.0000 EQ/L	7.9700 - 8.0300	7.9766
Sample ID			
Tested 12/12/24			
Tester ENOR			
Recorded 12/12/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer PO # :

Actual Density	1.27539 G/ML	1.27344 - 1.27733	1.27475
Sample ID			
Tested	12/12/24		
Tester	ENOR		
Recorded	12/12/24		
N.I.S.T. Traceable to SRM 84	YES YES		YES
Sample ID			
Tested	12/12/24		
Tester	ENOR		
Recorded	12/12/24		
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		12/30/26
Sample ID			
Tested	12/12/24		
Tester	ENOR		
Recorded	12/12/24		
DATE OF SAMPLING FOR TEST	MM/DD/YYYY		12/12/2024
PURPOSES	MM/DD/YYYY		
Sample ID			
Tested	12/12/24		
Tester	ENOR		
Recorded	12/12/24		
Lot #	4112121		
Made	12/12/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager