12/06/24

Made

Page: 1

Date: 12/06/24 at 11:36 AM

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

uonago	d Product:		STANDARD 100.0 PPM W/W .	AOIIFOIIS	
Test		DILLON	Target/UOM	Range	Result
	D TO FORMU	LATION ON	YES	- italige	YES
FILE			YES		
	Sample ID				
	Tested	12/06/24			
	Tester	BPHI			
	Recorded	12/06/24			
APPEARANCE AND COLO			CLEAR WATER WHITE		PASS
		-	LIQUID / NONE		
			APPEARANCE		
	Sample ID				
	Tested	12/06/24			
	Tester	BPHI			
	Recorded	12/06/24			
INSTRUMENTS USED DURING			LAB EQUIPMENT		P176,VF1-Z16
PREPARATION			INSTRUMENT		-,
	Sample ID				
	Tested	12/06/24			
	Tester	BPHI			
	Recorded	12/06/24			
EXPIRATION DATE			MM/DD/YYYY		04/30/2025
			MM/DD/YYYY		
	Sample ID				
	Tested	12/06/24			
	Tester	BPHI			
	Recorded	12/06/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager