



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4099			
Hydrochloric Acid 0.025 Normal NIST Traceable solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID			
Tested 12/20/24			
Tester JRIO			
Recorded 12/20/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES		YES
Sample ID			
Tested 12/20/24			
Tester JRIO			
Recorded 12/20/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY		12/20/2024
Sample ID	MM/DD/YYYY		
Tested 12/20/24			
Tester JRIO			
Recorded 12/20/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0	10.0 - 35.0	20.0
Sample ID	DEG C		
Tested 12/20/24			
Tester JRIO			
Recorded 12/20/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY		12/20/2024
Sample ID	MM/DD/YYYY		
Tested 12/20/24			
Tester JRIO			
Recorded 12/20/24			
AMBIENT TEST TEMPERATURE DEGC	20.0	15.0 - 30.0	20.0
Sample ID	DEG C		
Tested 12/20/24			
Tester JRIO			
Recorded 12/20/24			
NORMALITY TO 4 DEC PLACES	0.0250	0.0245 - 0.0255	0.0250
Sample ID	EQUIV/LITER		
Tested 12/20/24			
Tester JRIO			
Recorded 12/20/24			
Actual Density	0.99869	0.99858 - 0.99880	0.99876
Sample ID	G/ML		
Tested 12/20/24			
Tester JRIO			
Recorded 12/20/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD	YES
Sample ID		
Tested 12/20/24		
Tester JRIO		
Recorded 12/20/24		
N.I.S.T. TRACEABLE TO SRM 723	YES YES	YES
Sample ID		
Tested 12/20/24		
Tester JRIO		
Recorded 12/20/24		
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested 12/20/24		
Tester JRIO		
Recorded 12/20/24		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	VF3-J7,T-579-2895
Sample ID		
Tested 12/20/24		
Tester JRIO		
Recorded 12/20/24		
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY	12/30/2025
Sample ID		
Tested 12/20/24		
Tester JRIO		
Recorded 12/20/24		
Lot # 4120512		
Made 12/20/24		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.



Quality Manager