

Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO # :

Packaged Produ	ct: EQS523	3F		
. denuged i ledu		TANDARD 20.0 PPM (CELI	TE) SHAKE WEL	
Test		Target/UOM	Range	Result
PREPARED TO FOR		YES	Kange	YES
FILE		YES		120
Sample	חו	120		
Tested	12/19/24			
Tester	AROS			
Recorde				
PPM TOTAL SUSPEI		21.0	20.0 - 22.0	21.1
		PPM		
Sample	חו			
Tested	12/19/24			
Tester	AROS			
Recorde				
APPEARANCE AND COLOR		CLEAR TO SLIGHTLY TURBID		PASS
		LIQUID		
		APPEARANCE		
Sample	ID			
Tested	12/19/24			
Tester	AROS			
Recorde				
INSTRUMENTS USED DURING		INSTRUMENT		M108,VF4-N19
PREPARATION				
Sample	ID			
Tested	12/19/24			
Tester	AROS			
Recorde	ed 12/19/24			
EXPIRATION DATE		MM/DD/YYYY		12/30/2025
		MM/DD/YYYY		
Sample	ID			
Tested	12/19/24			
Tester	AROS			
Recorde	ed 12/19/24			
Lot # 412	21845			
	19/24			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

1