



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 4112			
HYDROCHLORIC ACID 0.02 NORMAL IN METHANOL			
Test	Target/UOM	Range	Result
Actual Normality	0.0200 EQ/L	0.0195 - 0.0205	0.0195
Sample ID Tested 12/19/24 Tester ENOR Recorded 12/19/24			
SAMPLE COLLECTION & HANDLING PER LAB040	YES YES		YES
Sample ID Tested 12/19/24 Tester ENOR Recorded 12/19/24			
DATE OF SAMPLING FOR TEST PURPOSES	MM/DD/YYYY MM/DD/YYYY		12/19/2024
Sample ID Tested 12/19/24 Tester ENOR Recorded 12/19/24			
AMBIENT TEMPERATURE DURING SAMPLING DEGC	25.0 DEG C	10.0 - 35.0	19.3
Sample ID Tested 12/19/24 Tester ENOR Recorded 12/19/24			
DATE OF TESTING OF SAMPLE	MM/DD/YYYY MM/DD/YYYY		12/19/2024
Sample ID Tested 12/19/24 Tester ENOR Recorded 12/19/24			
AMBIENT TEST TEMPERATURE DEG C	20.0 DEG C	10.0 - 30.0	20.0
Sample ID Tested 12/19/24 Tester ENOR Recorded 12/19/24			
QCSP5: STANDARDIZATION OF AQUEOUS ACIDS	YES METHOD		YES
Sample ID Tested 12/19/24 Tester ENOR Recorded 12/19/24			
Actual Density	0.79274 G/ML	0.79221 - 0.79328	0.79281
Sample ID Tested 12/19/24 Tester ENOR Recorded 12/19/24			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

N.I.S.T. Traceable to SRM 723	Yes YES	YES
Sample ID		
Tested	12/19/24	
Tester	ENOR	
Recorded	12/19/24	
PREPARED TO FORMULATION ON FILE	YES YES	YES
Sample ID		
Tested	12/19/24	
Tester	ENOR	
Recorded	12/19/24	
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE	PASS
Sample ID		
Tested	12/19/24	
Tester	ENOR	
Recorded	12/19/24	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M112,VF2-B16
Sample ID		
Tested	12/19/24	
Tester	ENOR	
Recorded	12/19/24	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	03/30/25
Sample ID		
Tested	12/19/24	
Tester	ENOR	
Recorded	12/19/24	
SHELF LIFE OF PRODUCT	3 MONTHS TIME	3 MONTHS
Sample ID		
Tested	12/19/24	
Tester	ENOR	
Recorded	12/19/24	
Lot #	4121363	
Made	12/19/24	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 3

Date: 12/19/24 at 3:12 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager