



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: FIS112F Phosphate Standard 1.0 ppm Solution			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 12/18/24			
Tester MMAT			
Recorded 12/18/24			
CALCULATED PPM PHOSPHATE (PO4)	1.00 PPM (MG/L)	0.98 - 1.02	1.00
Sample ID			
Tested 12/18/24			
Tester MMAT			
Recorded 12/18/24			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 12/18/24			
Tester MMAT			
Recorded 12/18/24			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P137,VF1-Z25
Sample ID			
Tested 12/18/24			
Tester MMAT			
Recorded 12/18/24			
EXPIRATION DATE	MM/DD/YYYY MM/DD/YYYY		06/30/2025
Sample ID			
Tested 12/18/24			
Tester MMAT			
Recorded 12/18/24			
Lot #	4121871		
Made	12/18/24		

Richard [Signature]
Quality Manager