12/17/24

Made

Page: 1

Date: 12/17/24 at 6:49 PM

## Aqua Solutions, Inc. 6913 Highway 225 Deer Park, TX, 77536 281/479-2569 Certificate of Analysis / QC Results

Customer PO #:

Packaged Product	: 209827F	,		
· ·	VERSENA	TE HARDNESS TITRANT (0.	01 MOLAR) AF GRADE	
Test		Target/UOM	Range	Result
ACTUAL MOLARITY		0.0100	0.0097 - 0.0103	0.0100
		MOLARITY (M/L)		
Sample ID				
Tested	12/17/24			
Tester	JRIO			
Recorded	12/17/24			
APPEARANCE AND COLOR		CLEAR LIQUID		PASS
		APPEARANCE		
Sample ID				
Tested	12/17/24			
Tester	JRIO			
Recorded	12/17/24			
INSTRUMENTS USED DURING		INSTRUMENT		M112,GC,VF5-L17,M55,BURET
PREPARATION				:079
Sample ID				
Tested	12/17/24			
Tester	JRIO			
Recorded	12/17/24			
EXPIRATION DATE		MM/DD/YY		12/30/26
		MM/DD/YYYY		
Sample ID				
Tested	12/17/24			
Tester	JRIO			
Recorded	12/17/24			
Lot # 4120	574			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager** 

hichard Stot